



| REVISIONS | | | |
|-----------|-------------|------|----------|
| REV | DESCRIPTION | DATE | APPROVED |

- 1 ISO: NW50 KF Flange
- 2 4-40 UNC x .160 deep 8 places



VACUUM SIDE



Accu-Glass Products Inc.
700 Arroyo Ave.
San Fernando, CA. 91340
TEL. 818-365-4215
FAX. 818-365-7074

Notes :

*** UNCONTROLLED WEB COPY ***

| | |
|---------------------------------------|---------|
| Material 304 SS | |
| Tolerances unless otherwise specified | |
| Decimal | Angular |
| .xx= +/- .005 | +/- 1° |
| .xxx= +/- .002 | |

| | | | |
|--|----------|---------|-------------------|
| Title 9D2-K50 | | | |
| 2 x 9-Pin Sub-D Feedthroughs on a NW50 KF Flange | | | |
| Drawn By | M.Wilson | DWG NO. | 100239 |
| Approved | | | Date 6-18-97 |
| | | | Cage Code : 4HCQ8 |