2

4

Date

Initial Release А

Rev Description

А

В

С

D

Factory calibrated for 2 PSI (lower pressure calibration available on request)
Leak tested to 5x10E-10 Standard cc/sec of He


1

| Flow Rate (SCFM) | * PRV Inlet Pressure (PSIG) | | | | |
|------------------|-----------------------------|--|--|--|--|
| 6.21 | 2.32 | | | | |
| 9.37 | 2.61 | | | | |
| 15.83 | 3.53 | | | | |
| 20.63 | 4.18 | | | | |
| 25.95 | 5.11 | | | | |
| 31.40 | 6.14 | | | | |
| | | | | | |



3

*Inlet pressure calculated at Albuquerque, NM altitude (12.1 psia)



| | Accu-Glass Products Inc.25047 Anza DriveProprietary and ConfideValencia, CA 91355The information containeT 818-365-4215property of Accu-Glass,F 818-365-7074in part or as whole, withe | UNCONTROLLED WEB DRAWING | Material | Componente | Title | | | | |
|--|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------|------------------------------------------|----------------------------------|----------------------------------|-----|--------|-----|-----------|
| | | Proprietary and Confidential The information contained in this drawing is the sole property of Accu-Glass, Products Inc. Any reproduction, | See Individual Components Tolerances. | | | | | | |
| | | | unless otherw | vise specified | Drawn by Ligeti Date 5/4/2016 | No. | 113155 | | Rev. A |
| | | in part or as whole, without the written permission of Accu-Glass Products, Inc., is strictly prohibited. | .XX ± .010 .XXX ± .005 | Angles \pm 0.5° Finish 63Ra | Approved | | | She | et |