


Notes

1. Leak test to 5x10E-10 Standard cc/sec of He.



Rev	Description	Date	Approved
A	Initial Release		


UV	IR	Graded Index	Fiber Size	Flange
-	-	112342	62.5µm	CF275
112426	112555	-	100µm	CF275
112300	112308	-	200µm	CF275
112301	112309	-	400µm	CF275
112302	112310	-	600µm	CF275
112303	112311	-	1000µm	CF275

2.75 CF Flange
304 Stainless Steel

Ø2.730

2x TIG

6x Ø.265
on Ø2.312 BC



Accu-Glass Products Inc.
25047 Anza Drive
Valencia, CA 91355
T 818-365-4215
F 818-365-7074
www.accuglassproducts.com

UNCONTROLLED WEB DRAWING

Proprietary and Confidential
The information contained in this drawing is the sole property of Accu-Glass, Products Inc. Any reproduction, in part or as whole, without the written permission of Accu-Glass Products, Inc., is strictly prohibited.

Material

Tolerances,
unless otherwise specified

.X ± .030
.XX ± .010
.XXX ± .005

Angles ± 0.5°
Finish 63Ra

Title

FO Dual Feedthrough on 2.75 CF Flange

Drawn by

Ligeti

Date

10/30/2013

No.

CF Dual Fiber VFT

Rev.

A

Approved

Sheet