




## Notes

1. Reducer Tee, 1.50~0.75 Tube / 2.75~1.33 CF Flange
2. Leak test to 5x10E-10 Standard cc/sec of He
3. Protect flange seal surfaces (knife-edge) from damage
4. Dimensions symmetrical about centerlines

Rev.	Description	Date	Made By	Approval
A	Initial Release			



**Accu-Glass Products Inc.**  
25047 Anza Drive  
Valencia, CA 91355  
T 818-365-4215  
F 818-365-7074  
www.accuglassproducts.com

**UNCONTROLLED WEB DRAWING**

Proprietary and Confidential  
The information contained in this drawing is the sole property of Accu-Glass, Products Inc. Any reproduction, in part or as whole, without the written permission of Accu-Glass Products, Inc., is strictly prohibited.

Material Stainless Steel, 304	
Tolerances, unless otherwise specified	
.X ± .030	Angles ± 0.5°
.XX ± .010	Finish 63Ra
.XXX ± .005	

Title Reducer Tee, 1.50~0.75 tube / 2.75~1.33 CF Flange			
Drawn by Date	Ligeti 11/5/2018	No. <b>200618</b>	Rev. <b>A</b>
Approved			Sheet